

norm

NEN-ISO 6297

Aardolieproducten - Vliegtuigbrandstoffen
en destillaten - Bepaling van de
elektrische geleiding (ISO 6297:1997)

Petroleum products - Aviation and distillate fuels - Determination of
electrical conductivity (ISO 6297:1997)

januari 1998
ICS 75.160.20

Vervangt NEN-ISO 6297:1984

Als Nederlandse norm is aanvaard:

- ISO 6297:1997

Normcommissie 310 028 "Aardolie en aardolieproducten"

Behoudens uitzondering door de wet gesteld mag zonder schriftelijke toestemming van het Nederlands Normalisatie-instituut niets uit deze uitgave worden verveelvoudigd en/of openbaar gemaakt door middel van fotokopie, microfilm, opslag in computerbestanden of anderszins, hetgeen ook van toepassing is op gehele of gedeeltelijke bewerking.

Het Nederlands Normalisatie-instituut is met uitsluiting van ieder ander gerechtigd de door derden verschuldigde vergoedingen voor verveelvoudiging te innen en/of daartoe in en buiten rechte op te treden, voor zover deze bevoegdheid niet is overgedragen c.q. rechtens toekomt aan de Stichting Reprorecht.

Hoewel bij deze uitgave de uiterste zorg is nagestreefd, kunnen fouten en onvolledigheden niet geheel worden uitgesloten. Het Nederlands Normalisatie-instituut en/of de leden van de commissies aanvaarden derhalve geen enkele aansprakelijkheid, ook niet voor directe of indirecte schade, ontstaan door of verband houdende met toepassing van door het Nederlands Normalisatie-instituut gepubliceerde uitgaven.

Dit document mag slechts op een stand-alone PC worden geïnstalleerd. Gebruik op een netwerk is alleen toegestaan als een aanvullende licentieovereenkomst voor netwerkgebruik met NEN is afgesloten. This document may only be used on a stand-alone PC. Use in a network is only permitted when a supplementary license agreement for use in a network with NEN has been concluded.

Preview

**Petroleum products — Aviation and
distillate fuels — Determination of electrical
conductivity**

*Produits pétroliers — Carburants aviation et distillats — Détermination
de la conductivité électrique*

Preview



Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 6297 was prepared by Technical Committee ISO/TC 28, *Petroleum products and lubricants*.

This second edition cancels and replaces the first edition (ISO 6297:1983), which has been technically revised.

Annex A of this International Standard is for information only.

© ISO 1997

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet central@iso.ch
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland

Petroleum products — Aviation and distillate fuels — Determination of electrical conductivity

WARNING — The use of this International Standard may involve hazardous materials, operations and equipment. This International Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this International Standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

1 Scope

This International Standard specifies a method for the determination of the electrical rest conductivity of aviation and distillate fuels with or without a static dissipator additive. The method measures conductivity when the fuel is uncharged, i.e. electrically at rest (known as rest conductivity).

Two methods are available for fuel conductivity measurement, either using a portable meter for the direct measurement in tanks or laboratory and field measurement of samples, or using an in-line meter for continuous measurement of fuel conductivities (equivalent to rest conductivity) in a flowing stream. When using either type of instrument care should be taken in allowing the relaxation of residual electrical charges before measurement and in preventing contamination.

The procedure can be used to measure conductivities from 1 pS/m to 2 000 pS/m using the equipment used to establish the reported precision. The range can be extended to greater than 2 000 pS/m with the correct selection of electrode size and current measuring apparatus.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent edition of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 3171:1988, *Petroleum liquids - Automatic pipeline sampling.*

ASTM D4057 : 1988, *Standard Practice for Manual Sampling of Petroleum and Petroleum Products*

3 Definitions

For the purposes of this International Standard the following definition applies.

3.1 rest conductivity: The reciprocal of the resistivity of uncharged fuel in the absence

of ionic depletion or polarization.

This is the electrical conductivity, expressed in picosiemens per metre (pS/m), at the initial instant of current measurement after a d.c. voltage is impressed between the electrodes.

NOTE 1 — The term conductivity unit (CU) is equal to 1 pS/m. A siemens is the SI definition of reciprocal ohm sometimes called mho (Ω^{-1}).

4 Principle

A voltage is applied across two electrodes immersed in the fuel and the resulting current expressed as a conductivity value. With portable meters, the current measurement is made almost instantaneously upon application of the voltage to avoid errors arising from ion depletion. Ion depletion or polarization is eliminated in dynamic monitoring systems by continuous replacement of the sample in the measuring cell.

5 Reagents

5.1 Propan-2-ol, reagent grade.

5.2 Toluene, reagent grade.

6 Apparatus

6.1 Portable conductivity cell and current-measuring apparatus, capable of giving a conductivity reading almost instantaneously with the application of a voltage.

6.2 In-line conductivity measuring cell, incorporating a facility to remove electrostatic charges before the representative fuel stream is passed through an in-line measuring cell.

NOTE 2 — A controlled continuous flow through the fuel cell prevents ion depletion thereby providing the equivalent of rest conductivity as a continuous measurement. This in-line instrument is designed for permanent installation in the fuel distribution system. It includes a pump to ensure constant flow through the fuel cell and also a sensor to monitor temperature.

Any sample tapping point shall be installed at least 30m downstream of any additive injection system, unless a mixing device is used which has been shown to give adequate mixing of the additive concerned prior to sampling.

6.3 Temperature sensor, having a range suitable for measuring the temperature of fuel samples in the field to ± 1 °C. This is only required when using the portable conductivity cell.

6.4 Measuring vessel, capable of holding sufficient fuel sample to cover the electrodes of the conductivity measuring cell to the depth required for that particular piece of apparatus.

7 Sampling

To avoid errors in the measurement of fuel conductivity, either carry out the measurements *in situ*, or take samples using the procedures specified in ASTM D4057 or ISO 3171. The following additional precautions are also required:

- a) the sample size shall be as large as practicable and not less than 1 litre;
- b) containers shall be fully epoxy-lined cans or of polytetrafluoroethylene (PTFE) only (see note 3);
- c) prior to sampling, all containers and their closures shall be rinsed at least three times with the fuel being sampled;
- d) samples shall be tested as soon as possible after sampling.

NOTE 3 — Test method results are known to be sensitive to trace contamination during the sampling operation and from sample containers. New containers are recommended, but when only used containers are available they shall be thoroughly rinsed with cleaning solvent, propan-2-ol (5.1), followed by toluene (5.2), and dried with a stream of air.

8 Procedure

8.1 Calibration

Carry out calibration procedures in accordance with the equipment manufacturer's instructions immediately prior to use of the apparatus.

8.2 *In situ* field measurements in tanks, tank cars, trucks, etc.

NOTES

4 — For field measurements portable conductivity meters are considered suitable. In order to satisfy local safety regulations they should be certified as intrinsically safe for use in hazardous locations. Each meter has, or can be equipped with, an extension cable to lower the cell into the tank.

5 — High impedance hand-held meters are susceptible to electrical transients caused by the extension cable flexing during measurements. Failure to hold the apparatus steady may result in significantly reduced precision.

8.2.1 Check the calibration and earth (ground) the meter to the tank. Lower the conductivity cell into the tank to the desired level, taking care to avoid partial immersion or contact with tank water bottoms, if present. Move the conductivity cell in an up-and-down motion to remove air and previous fuel residues.

Bestelformulier

NEN

Stuur naar:

NEN Uitgeverij
t.a.v. afdeling Marketing
Antwoordnummer 10214
2600 WB Delft

NEN Uitgeverij

Postbus 5059
2600 GB Delft

Vlinderweg 6
2623 AX Delft

T (015) 2 690 390

F (015) 2 690 271

www.nen.nl/normshop

Ja, ik bestel

__ ex. NEN-ISO 6297:1998 en Aardolieproducten - Vliegtuigbrandstoffen en destillaten - Bepaling van de elektrische geleiding € 48.19

Wilt u deze norm in PDF-formaat? Deze bestelt u eenvoudig via www.nen.nl/normshop

Stel uw vraag aan
Klantenservice via:

[@NEN_webcare](https://twitter.com/NEN_webcare)

Gratis e-mailnieuwsbrieven

Wilt u op de hoogte blijven van de laatste ontwikkelingen op het gebied van normen, normalisatie en regelgeving? Neem dan een gratis abonnement op een van onze e-mailnieuwsbrieven. www.nen.nl/nieuwsbrieven

Retourneren

Fax: (015) 2 690 271
E-mail: marketing@nen.nl
Post: NEN Uitgeverij,
t.a.v. afdeling Marketing
Antwoordnummer 10214,
2600 WB Delft
(geen postzegel nodig).

Gegevens

Bedrijf / Instelling

T.a.v. _____ O M O V

E-mail

Klantnummer NEN

Uw ordernummer _____ BTW nummer _____

Postbus / Adres

Postcode _____ Plaats _____

Telefoon _____ Fax _____

Factuuradres (indien dit afwijkt van bovenstaand adres)

Postbus / Adres

Postcode _____ Plaats _____

Datum _____ Handtekening _____

Voorwaarden

- De prijzen zijn geldig tot 31 december 2015, tenzij anders aangegeven.
- Alle prijzen zijn excl. btw, verzend- en handelingskosten en onder voorbehoud bij o.m. ISO- en IEC-normen.
- Bestelt u via de normshop een pdf, dan betaalt u geen handeling en verzendkosten.
- Meer informatie: telefoon (015) 2 690 391, dagelijks van 8.30 tot 17.00 uur.
- Wijzigingen en typfouten in teksten en prijsinformatie voorbehouden.
- U kunt onze algemene voorwaarden terugvinden op: www.nen.nl/leveringsvoorwaarden.

Normalisatie: de wereld op één lijn.

preview - 2015