

Gas analysis - Preparation of calibration gas mixtures using dynamic volumetric methods - Part 1: Methods of calibration (ISO 6145-1:2003, IDT)

december 2003
ICS 71.040.40

Als Nederlandse norm is aanvaard:

- ISO 6145-1:2003, IDT

Normcommissie 310 158 "Gasanalyse"

Apart from exceptions provided by the law, nothing from this publication may be duplicated and/or published by means of photocopy, microfilm, storage in computer files or otherwise, which also applies to full or partial processing, without the written consent of the Netherlands Standardization Institute.

The Netherlands Standardization Institute shall, with the exclusion of any other beneficiary, collect payments owed by third parties for duplication and/or act in and out of law, where this authority is not transferred or falls by right to the Reproduction Rights Foundation.

Auteursrecht voorbehouden. Behoudens uitzondering door de wet gesteld mag zonder schriftelijke toestemming van het Nederlands Normalisatie-instituut niets uit deze uitgave worden veeleenvoudigd en/of openbaar gemaakt door middel van fotokopie, microfilm, opslag in computerbestanden of anderszins, hetgeen ook van toepassing is op gehele of gedeeltelijke bewerking.

Het Nederlands Normalisatie-instituut is met uitsluiting van ieder ander gerechtigd de door derden verschuldigde vergoedingen voor veeleenvoudiging te innen en/of daartoe in en buiten rechte op te treden, voor zover deze bevoegdheid niet is overgedragen c.q. rechs toekomt aan de Stichting Reprorecht.

Although the utmost care has been taken with this publication, errors and omissions cannot be entirely excluded. The Netherlands Standardization Institute and/or the members of the committees therefore accept no liability, not even for direct or indirect damage, occurring due to or in relation with the application of publications issued by the Netherlands Standardization Institute.

Hoewel bij deze uitgave de uiterste zorg is nagestreefd, kunnen fouten en onvolledigheden niet geheel worden uitgesloten. Het Nederlands Normalisatie-instituut en/of de leden van de commissies aanvaarden derhalve geen enkele aansprakelijkheid, ook niet voor directe of indirecte schade, ontstaan door of verband houdend met toepassing van door het Nederlands Normalisatie-instituut gepubliceerde uitgaven.

Dit document mag slechts op een stand-alone PC worden geïnstalleerd. Gebruik op een netwerk is alleen toegestaan als een aanvullende licentieovereenkomst voor netwerkgebruik met NEN is afgesloten. This document may only be used on a stand-alone PC. Use in a network is only permitted when a supplementary license agreement for use in a network with NEN has been concluded.

Preview

Voorbeeld
Preview

INTERNATIONAL STANDARD

ISO
6145-1

Second edition
2003-11-15

Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods —

Part 1: Methods of calibration

*Analyse des gaz — Préparation des mélanges de gaz pour étalonnage
à l'aide de méthodes volumétriques dynamiques —*

Partie 1: Méthodes d'étalonnage

Preview



Reference number
ISO 6145-1:2003(E)

© ISO 2003

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

Copyright
Preview

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Calibration methods.....	2
4.1 General.....	2
4.2 Description of primary or potentially primary measuring devices.....	4
4.3 Measurements on the final mixture.....	12
5 Techniques for preparation of gas mixtures calibrated by the methods described in Clause 4.....	13
5.1 General.....	13
5.2 Volumetric pumps (see ISO 6145-2 ^[3]).....	15
5.3 Continuous injection (see ISO 6145-4 ^[4]).....	15
5.4 Capillary (see ISO 6145-5 ^[5]).....	15
5.5 Critical orifices (see ISO 6145-6 ^[6]).....	16
5.6 Thermal mass flow controllers (see ISO 6145-7 ^[7]).....	16
5.7 Diffusion (see ISO 6145-8 ^[8]).....	16
5.8 Saturation (see ISO 6145-9 ^[9]).....	17
5.9 Permeation (see ISO 6145-10 ^[10]).....	17
Annex A (normative) Volume measurement by weighing the water content.....	19
Annex B (informative) Description of secondary devices which need calibration against primary devices.....	23
Bibliography.....	32

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 6145-1 was prepared by Technical Committee ISO/TC 158, *Analysis of gases*.

This second edition cancels and replaces the first edition (ISO 6145-1:1986), in which the estimated uncertainties in the calibration methods and techniques have now been combined in a square-root sum-of-squares manner to form the relative combined standard uncertainty. In comparison with the previous edition the periodic injection has been deleted (limited application).

ISO 6145 consists of the following parts, under the general title *Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods*:

- *Part 1: Methods of calibration*
- *Part 2: Volumetric pumps*
- *Part 4: Continuous injection methods*
- *Part 5: Capillary calibration devices*
- *Part 6: Critical orifices*
- *Part 7: Thermal mass-flow controllers*
- *Part 9: Saturation method*
- *Part 10: Permeation method*

Diffusion will be the subject of a future Part 8 to ISO 6145. Part 3 to ISO 6145, entitled *Periodic injections into a flowing gas*, has been withdrawn.

Introduction

This part of ISO 6145 is one of a series of standards which describes the various dynamic volumetric methods used for the preparation of calibration gas mixtures.

In dynamic volumetric methods a gas, A, is introduced at volume or mass flow rate q_A into a constant flow rate q_B of a complementary gas B. Gas A can be either a pure calibration component, i , or a mixture of i in A.

The volume fraction, $\varphi_{i,M}$ of i in the final calibration gas mixture is given in the following equation:

$$\varphi_{i,M} = \varphi_{i,A} \left(\frac{q_A}{q_A + q_B} \right)$$

where $\varphi_{i,A}$ is the volume or mass fraction of component, i , in the pre-mixed gas A, and is already known from its method of preparation. It is assumed that in this equation, $\varphi_{i,B}$, the concentration of component, i , in gas B, is zero.

The introduction of gas A can be continuous (e.g. permeation tube) or pseudo-continuous (e.g. volumetric pump). A mixing chamber should be inserted in the system before the analyser and is particularly essential in the case of pseudo-continuous introduction. The flow rate of component A is measured either directly in terms of volume or mass, or indirectly by measuring the variation of a physical property.

The dynamic volumetric preparation techniques produce a continuous flow rate of calibration gas mixtures into the analyser but do not generally allow the build-up of a reserve by storage under pressure.

The main techniques used for the preparation of the mixtures are:

- a) volumetric pumps;
- b) continuous injection;
- c) capillary;
- d) critical orifices;
- e) thermal mass-flow controllers;
- f) diffusion;
- g) saturation;
- h) permeation;
- i) electrochemical generation.

In all cases, and most particularly if very dilute mixtures are concerned, the materials used for the apparatus are chosen as a function of their resistance to corrosion and low absorption capacity (usually glass, PTFE or stainless steel). It should, however, be pointed out that the phenomena are less important for dynamic volumetric methods than for static methods.

Numerous variants or combinations of the main techniques can be considered and mixtures of several constituents can also be prepared by successive operations.

ISO 6145-1:2003(E)

Some of these techniques allow calculation of the final concentration of the gas mixture from basic physical information (e.g. mass rates of diffusion, flow through capillaries). However, since all techniques are dynamic and rely on stable flow rates, this part of ISO 6145 emphasizes calibration of the techniques by measurement of the individual flow rates or their ratios, or by determination of the composition of the final mixture.

The uncertainty of the composition of the calibration gas mixture is best determined by comparison with a gas mixture traceable to international standards. Certain of the techniques which may be used to prepare a range of calibration gas mixtures may require several such traceable gas mixtures to verify their performance over that range. The dynamic volumetric technique used has a level of uncertainty associated with it. Information on the final mixture composition depends both on the calibration method and on the preparation technique.

Voorbereid
Preview

Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods —

Part 1: Methods of calibration

1 Scope

This part of ISO 6145 specifies the calibration methods involved in the preparation of gas mixtures by dynamic volumetric techniques. It also gives a brief presentation of a non-exhaustive list of examples of dynamic volumetric techniques which are described in more detail in other parts of ISO 6145.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6142, *Gas analysis — Preparation of calibration gas mixtures — Gravimetric method*

ISO 6143, *Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures*

ISO 7504, *Gas analysis — Vocabulary*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 7504 and the following apply.

3.1

uncertainty of measurement

parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the measurand

NOTE 1 Values of the individual statistical uncertainties found in some methods and techniques in this part of ISO 6145 are combined with the values of systematic uncertainties that also occur in a square-root sum-of-squares manner to provide a relative combined uncertainty, or in some cases as a relative expanded uncertainty by application of the coverage factor "2".

NOTE 2 In keeping with Reference [1] of the Bibliography, the uncertainty of the composition of a mixture is expressed as a relative expanded uncertainty.

Bestelformulier

Stuur naar:

NEN Standards Products & Services
t.a.v. afdeling Klantenservice
Antwoordnummer 10214
2600 WB Delft



NEN Standards Products & Services

Postbus 5059
2600 GB Delft

Vlinderweg 6
2623 AX Delft

T (015) 2 690 390
F (015) 2 690 271

www.nen.nl/normshop

Ja, ik bestel

___ ex. NEN-ISO 6145-1:2003 en Gasanalyse - Bereiding van kalibratiegasmengsels met gebruik van dynamische volumetrische methoden - Deel 1: Kalibratiemethoden

€ 125.90

Wilt u deze norm in PDF-formaat? Deze bestelt u eenvoudig via www.nen.nl/normshop

Gratis e-mailnieuwsbrieven

Wilt u op de hoogte blijven van de laatste ontwikkelingen op het gebied van normen, normalisatie en regelgeving? Neem dan een gratis abonnement op een van onze e-mailnieuwsbrieven. www.nen.nl/nieuwsbrieven

Gegevens

Bedrijf / Instelling _____

T.a.v. _____ O M O V

E-mail _____

Klantnummer NEN _____

Uw ordernummer _____ BTW nummer _____

Postbus / Adres _____

Postcode _____ Plaats _____

Telefoon _____ Fax _____

Factuuradres (indien dit afwijkt van bovenstaand adres)

Postbus / Adres _____

Postcode _____ Plaats _____

Datum _____ Handtekening _____

Retourneren

Fax: 015 2 690 271

E-mail: klantenservice@nen.nl

Post: NEN Standards Products & Services,

t.a.v. afdeling Klantenservice
Antwoordnummer 10214,
2600 WB Delft

(geen postzegel nodig).

Voorwaarden

- De prijzen zijn geldig tot 31 december 2018, tenzij anders aangegeven.
- Alle prijzen zijn excl. btw, verzend- en handelingskosten en onder voorbehoud bij o.m. ISO- en IEC-normen.
- Bestelt u via de normshop een pdf, dan betaalt u geen handeling en verzendkosten.
- Meer informatie: telefoon 015 2 690 391, dagelijks van 8.30 tot 17.00 uur.
- Wijzigingen en typfouten in teksten en prijsinformatie voorbehouden.
- U kunt onze algemene voorwaarden terugvinden op: www.nen.nl/leveringsvoorwaarden.